



**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

 MANUFACTURER:	CONTEC MEDICAL SYSTEMS CO., LTD No.112 Qinhuang West Street, Economic & Technical Development Zone, Qinhuangdao, Hebei Province, PEOPLE'S REPUBLIC OF CHINA		
MEDICAL DEVICE:	Pulse Oximeter Probe, ESC0029		
CLASSIFICATION - ANNEX IX:	Class II b, Rule 1		
CONFORMITY ASSESSMENT ROUTE:	Annex II excluding chapter 4		
WE, (CONTEC MEDICAL SYSTEMS CO., LTD) HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES. MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURE.			
STANDARDS APPLIED: SEE ATTACHED LIST OF (HARMONISED - EN) STANDARDS FOR WHICH DOCUMENTED EVIDENCE OF COMPLIANCE CAN BE PROVIDED.			
NOTIFIED BODY:	TÜV SÜD PRODUCT SERVICE GMBH RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY		
IDENTIFICATION NUMBER:	CE 0123		
(EC) CERTIFICATE(S):	<u>G1 050972 0050 Rev.04</u>		
<table border="1"><tr><td>EC</td><td>REP</td></tr></table>	EC	REP	Shanghai International Holding Corp. GmbH(Europe) Eiffestrasse 80, 20537 Hamburg Germany
EC	REP		
EUROPEAN REPRESENTATIVE:			

START OF CE-MARKING: 2009-07-23 (Date or Lot or serial number)

PLACE, DATE OF DECLARATION:	QINHUANGDAO, 2020-06-18
SIGNATURE:	 President

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC
CONCERNING MEDICAL DEVICES**

Appendix: list of (harmonised - EN) standards

NO.	Reference	Title
1	EN60601-1:1990+A1:1993+A2:1995 (IEC60601-1:1988+A1:1991+A2:1995)	Medical electrical equipment- Part 1: General requirements for safety
2	EN 60601-1-2:2007 (IEC60601-1-2:2007)	Medical electrical equipment- Part 1-2: General requirements for safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests
3	EN ISO 9919:2009	Medical electrical equipment - Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use