

EU DECLARATION OF CONFORMITY

This is a declaration made in accordance with the requirements of the following relevant Union harmonisation legislation. The manufacturer assures that the device that is covered by the present declaration is in conformity with this Regulation (EU) 2017/745 for Medical Devices and, if applicable, with any other relevant Union legislation that provides for the issuing of an EU declaration of conformity. The declaration of conformity is issued under the sole responsibility of the manufacturer.



Manufacturer's Name: SHANGHAI KOHDEN MEDICAL ELECTRONIC INSTRUMENT CORP.
Address: No. 567 Huancheng Bei Road, Shanghai Comprehensive Industrial Development Zone, Fengxian District, Shanghai, 201401, China
SRN: CN-MF-000013567
European Representative: NIHON KOHDEN EUROPE GmbH
Address: Raiffeisenstrasse 10, 61191 Rosbach, Germany
SRN: DE-AR-000010740

■ **Regulation (EU) 2017/745(MDR)**

Classification/Risk Class: IIa

Conformity assessment procedure:

Annex IX Chapter I and III

Notified Body

Name and No. : BSI Group The Netherlands B.V.; 2797
EC Certificate: MDR 748567

■ **Directive 2011/65/EU and 2015/863/EU**

Standard Applied: EN 50581:2012

□ **Directive 2014/53/EU (RED)**

Notified Body

Name and No. :
EU-Type Examination Certificate No. :
Standard Applied:

Declaration No.: SKXY-10008

Product Name, Model Number and Basic UDI-DI :

Product Name	Model Number	Basic UDI-DI	MDR	RoHS	RED
Electrocardiograph	ECG-2250	06970010880357	√	√	×
		06970010880364	√	√	×
		06970010880371	√	√	×
		06970010880388	√	√	×
Software kit	QS-041E	06970010880876	√	√	×

Intended purpose: The Shanghai Kohden ECG-2250 electrocardiograph is intended for medical use to process the electrical signals produced by the heart, which are acquired through two or more electrodes, and to display waveforms and/or prepare a record of these electrical signals. This device is a portable ECG acquisition terminal which measures up to 12-lead ECG waveforms. The intended use is only for diagnosis, not for monitoring of vital physiological parameters.
ECG-2250 is intended to be used by qualified medical personnel within a medical facility, such as hospital or clinic and children and adults can be diagnosed with this electrocardiograph.
The software kit QS-041E is the software for ECG-2250. It is intended for upgrading.

Additional Information:

Along with different optional accessories, OP No.100D, 101D, 102D, 103D are provided.

Authorized Signatory:
Shanghai, China/ 2022.12.2
Place and date of issue

Jiaoying Jin
General Manager
Quality Management Division